

Certificate of Calibration

Certificate Number: XXXXX
Certificate Date: XXXXXXXXX
Customer: XXXXXXXXX

Instrument Under Test: Model No EL-USB-1-LCD **Instrument Serial Number:** 000040780

Test Equipment Used: Climatic Chamber **Test Equipment Serial Number:** ID0896
 Rotronic Hygroclip 2 61396959
 HW4 24 00004727 49128

The Hygroclip 2 is certified to 'traceable' national standards.
 The Climatic Chamber is used to generate the temperature and humidity source.

Lab Environmental conditions at time of calibration: Temp 20.5°C±1°C/ RH 52% RH ±5% RH

| True Reading (°C) | Instrument Measured Value (°C) | | Error (°C) | | Uncertainty of Measurement (°C) |
|----------------------|-----------------------------------|-------|---------------|-------|---------------------------------------|
| | DISPLAY | GRAPH | DISPLAY | GRAPH | |
| -9.94 | -10 | -10.5 | 0.06 | 0.56 | ± (0.15°C + D) |
| 5.07 | 4.5 | 5.0 | 0.57 | 0.07 | ± (0.15°C + D) |
| 25.07 | 24.5 | 25.0 | 0.57 | 0.07 | ± (0.15°C + D) |
| 40.08 | 39.5 | 39.5 | 0.58 | 0.58 | ± (0.15°C + D) |

Temperature cycle and uniformity was conducted at full immersion with readings observed 5 minutes after stabilisation.

The reported expanded uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%.

Page 2 of 2 Pages

Certificate No. XXXXX

Qualification of Standard Used

This is to certify that the stated instrument has been verified and calibrated at the measured values given in the results table on page 1 of this 2 page document using test equipment which itself has been calibrated and certified to the stated standard. The procedures and work instructions used to verify and calibrate the instrument are fully documented.

The errors reported refer to measured values only with no account being taken of the instrument's ability to maintain its calibration.

Signed:

Authorised Signatory

Sean Wigmore